9·F·101

TRIPLETT

INSTRUCTION MANUAL

MODEL 3490

TRANSISTOR ANALYZER

MP MASTER

MA26153

PRICE $1.00

TABLE OF CONTENTS

SUBJECTS	Page
PHOTO, MODEL 3490	2
Meters and Range Selectors	3
The Main Controls	3
The Secondary Controls	4
Configuration—Test Index	5
Connections to Transistors	5
Caution	5
The Heat Sink	6
The Interlocks	6
The Warning Lights	6
Precautions	6
MEASUREMENT OR TEST PROCEDURE	
Test for Shorted Collector	7
COLLECTOR JUNCTION LEAKAGE CURRENT MEASUREMENTS	
Measuring I_{CO}	7
Measuring I_{CEO}	7
Measuring I_{CES}	7
Measuring I_{CER}	8
Measuring I_{CBR}	8
EMITTER JUNCTION LEAKAGE CURRENT MEASUREMENTS	
Measuring I_{EO} (I_{EBO})	8
Measuring I_{EBS} or I_{ECS}	8
Measuring I_{EBR}	8
THE MEASUREMENT OF ALPHA—h_{FB}	9
THE MEASUREMENT OF D. C. BETA—h_{FE}	9
MEASURING A. C. BETA—h_{fe}	10
MEASURING SATURATION VOLTAGE AND RESISTANCE	10
MEASURING PUNCH-THRU	10
Floating Potentials	11
TETRODE TRANSISTORS	11
TESTING DIODES AND RECTIFIERS	12
Testing Diodes	12
Testing Rectifiers	12
Selenium Rectifiers	13
Precaution When Testing All Semi-Conductors	13
BASIC TRANSISTOR TEST CIRCUITS	13
WIRING DIAGRAM, Model 3490	14
REPLACEABLE PARTS	15
WARRANTY	16

Copyright 1961 by

THE TRIPLETT ELECTRICAL INSTRUMENT CO.

Bluffton - - - - - Ohio


EMITTER or BASE CURRENT
Use switch multiplier
COLLECTOR CURRENT
Use switch multiplier
INPUT ON
COLL. ON
RET
12V
1.2V
INPUT VOLTS
PUNCH THRU
INPUT VOLTS
COLLECTOR VOLTS
PUNCH THRU
LABORATORY TRANSISTOR ANALYZER
60MA
6MA
600μA
60μA
SET B. 600
300
AC BETA
COLLECTOR
FUSE
ON
OFF
INPUT
SET AC BETA
TETRODE VOLTS
COLLECTOR VOLTS
MODEL 3490
PNP
NPN
COM. EM.
COM. BASE
SHORT
OPEN

Model 3490

TRIPLETT TRANSISTOR ANALYZER

MODEL 3490

This Transistor Analyzer is designed for maximum flexibility, wide coverage of both signal and power transistors, and standard readings of all measurement values in volts, amperes and ohms as listed by the semi-conductor manufacturers. Continuous current and continuous voltage controls are a built in feature, with large accurate meters to read static as well as dynamic values. Interlocked switching circuits with warning light indicators protect the semi-conductors from voltage transients, while at the same time assisting the operator in the testing procedure. Collector Current and Emitter or Base Current is monitored all the time, and the controls with other instrumentation provide facilities for plotting complete characteristic curves as well as taking single readings.

TECHNICAL DESCRIPTION

THE METERS AND RANGE SELECTORS:

Transistors are inherently current sensitive devices as compared to vacuum tubes which are primarily voltage sensitive. Current readings and current amplification values are therefore, the important guiding factors in transistor evaluation. It is of utmost importance to know the Collector Current at all times, regardless of the parameter under measurement. This Transistor Analyzer is equipped with a four and one half inch multirange ammeter which reads continuously the Collector Current on any one of 10 overlapping ranges. This meter is located in the upper right hand corner of the panel. The scale captioned "Collector Current" has two arcs; the outside one calibrated from zero to thirty, and the inside one from zero to ten. All of the current ranges are on a zero-ten and zero-thirty basis, from 1 milliampere to 30 amperes, and therefore, on any range **it is necessary only to add zero's** to the scale reading, or divide by ten to determine the exact Collector Current. As you will note, the high ampere range on the 3490 is 30 amperes. Care in operation of the Analyzer on this range is imperative. The 3490 is designed to handle over 100 D.C. volt-amperes of collector power. However, due to the heat characteristics which occur when this type of power is used, it is recommended that readings on this range should not exceed 90 seconds for any one reading.

When using the 3490 at the higher collector current values, it is recommended that for each 15 minutes continuous testing, it be allowed a similar period of cooling time to prevent a heat build-up in the components. The Collector Current range selector switch is mounted directly below the Collector Current meter providing in combination an effective and accurate milliammeter and ammeter accurate to ±2%. In addition to the current ranges, this switch can be set to the "Open" and "Short" positions, which are used when taking emitter readings under conditions where the collector must be open or short circuited to the base.

Since this Transistor Analyzer has been designed to measure three or four terminal units in either the Common Base or Common Emitter configuration, the Input Current can be introduced or applied thru either the Emitter or Base. Accordingly, this meter is labeled "Input Current", and is mounted in the upper left hand corner of the panel. It reads Emitter Current when the Common Emitter-Common Base lever switch is in the Common Base position, and when the switch is indexed to the Common Emitter position, the Input meter reads Base Current. It was stated before and is repeated here that transistors are current amplifying devices, and therefore, this meter is connected to read Input Current **all the time.** Like the Collector Current meter, the Input Current meter has the same two arcs numbered zero to thirty. Continuous coverage is available from 300 to ten and zero to thirty. Continuous coverage is available from 300 microamperes full scale to 1 ampere full scale, accurate to ±2%. Again it is necessary only to add zero's or divide by ten to obtain the exact emitter or base current applied to the transistor. Also, the Input Current range selector switch is mounted directly below the Input meter for operational convenience. In addition to the current ranges, this switch can also be set to "Open" and "Short" positions, which are used when taking collector readings under conditions where either the emitter or the base must be open or short circuited to the collector circuit. It was stated before that the Collector meter reads Collector Current at all times, and that the Input meter reads Input Current all of the time. Therefore, the operator of the Transistor Analyzer has in front of him the two most important functional values on continuous display. Thus, by comparing the Input or Emitter Current and the Collector Current in the Common Base configuration, the value of alpha or forward current transfer ratio h_{FB} can be determined. By comparing the ratio between Collector Current and the Input or Base Current in the Common Emitter configuration, the value of d-c Beta or forward current transfer ratio h_{FE} is determined.

The third meter in the approximate center of the panel reads Collector Voltage all the time unless it is being used for some special function, such as a-c Beta, I_{CO}, I_{CEO}, Input Voltage or Punch Thru. Six voltage ranges with about a 2-1 ratio are calibrated on the black or d-c arcs figured zero to 3, zero to 6, and zero to 12. For the 30, 60, and 120 volt ranges it is again necessary only to add a zero to the meter reading. These voltage ranges are available by using the Collector Volts range selector switch directly below the meter. The combination of the range switch and the Collector Volts meter gives the operator continuous control of the Collector voltage through 120 volts accurate to ±2%. The use of the three meters presents to the operator a completely flexible operational analysis of the transistor under test. Collector Current can be plotted as a function of Base Current by holding the Collector Voltage constant—Collector Current can be plotted against Collector Voltage—Saturation Voltage or resistance can be measured, etc.

The scale arcs figured in red on this meter are calibrated for the a-c Beta or h_{fe} readings, and are

used for this measurement on low power or signal transistor types. The range of a-c Beta is determined by the position of the a-c Beta rotary switch which is located at the lower right of this meter.

THE MAIN CONTROLS:

The main controls are set up on the panel to enable the operator to most conveniently explore the transistor characteristics and/or plot characteristic curves. There are three independent power supplies in the analyzer, namely, the collector supply with its control knob directly under the word "COLLECTOR", the input supply, with its control knob under the word "INPUT", and the tetrode supply with its control calibrated zero to 10 volts and marked "TETRODE VOLTS". The Collector control is a variable transformer controlling primary energy to the collector supply. The Input control is a vitreous power potentiometer for adjusting the secondary a-c voltage to the input supply. These are the two most important controls, and are equipped with the large knobs for convenience of adjustment. **They should always be returned to the extreme counter clockwise position before connecting a transistor for test.** The Tetrode control is used only for double base or tetrode transistors, and can be ignored except when testing these types.

THE SECONDARY CONTROLS:

Reading from left to right of the panel, you will see the "A. C. BETA" control. This is used in conjunction with the lower case "Beta Set" position of the "A. C. BETA" switch, and is to adjust the 1000 cycle injection current to the proper value, prior to measurement of A. C. Beta. This value is indicated by the "Set Beta" mark, which is above the top red arc on the center meter. This control, along with the other front panel components which are used for A. C. Beta measurements, are distinguished from the D. C. controls by the use of red figures on the panel and dial.

The INPUT current range switch and control have been previously discussed. These operate together for adjustment of the base or emitter current as desired.

Along the left side of the Collector Voltmeter is located a lever switch with the wording "RET", "Input Volts", and "Punch Thru". This is one of the three switches used for shifting circuits and for connecting the meter in the center of the panel into the required position. The "RET" is the panel abbreviation for "Return" and it is important that this switch be left in the "RET" position, except when taking the "Input Voltage" readings, or "Punch Thru" voltage readings, and then returned to the "RET" position before proceeding with any other tests. This lever switch may be indexed into the 12 volt input volts position, or into the 1.2 volt input position. Thus, in these two positions, the meter reads the input voltage directly at the emitter or base terminals from zero to 1.2 volts, or zero to 12 volts. If the Transistor Analyzer is set up for the common base configuration, the meter will read the voltage from the emitter to the base and collector return. Likewise, if it is set up in the common emitter configuration, the meter will read the voltage from the base to the emitter and collector return. In the "Punch Thru" position, the meter will indicate the voltage from the open emitter to the base, and as is marked on the panel, indicating voltage from zero to 1.2 volts on the Collector Voltmeter.

The lever switch along the right side of the Collector Voltmeter is used for all measurements of I_{CEO} and related leakage tests. This is the second of the three circuit shifting switches. You will note the "RET" position on this switch, which is, as on the Input Volts switch, in the uppermost position. This switch also **must** be indexed to the "RET" position except when taking I_{CO}, I_{CEO}, and related leakage tests, and immediately returned to the "RET" position. Since this analyzer handles power transistors up to 15 amperes collector current, as well as the lowest current signal types, wide coverage of these two parameters is required. Four current ranges are available, all of which are indicated on the Collector Voltmeter. The 60 ma, 6 ma, and 600 ua are indexed directly by the lever switch. For the last decade, the lever switch is left in the 600 ua position and the 60 ua push button switch is depressed. With the four decades of current, any value from one microampere or less to sixty milliamperes can be measured. The I_{CO} readings are of course taken in the common base configuration and I_{CEO} measurements in the common emitter configuration. This will be further explained under the "Test Procedure" section of these instructions. Since this is a progressive switch, and an unknown quantity is under measurement, the lever may be moved one step at a time until a suitable reading has been taken. If the indication on the meter is less than ten percent of full scale value, the lever switch may be indexed to the next decade down. This observation will prevent overloading of the current range and thus prevent pointer and meter damage.

Directly below this lever switch, and to the lower right of the Collector Voltmeter, is the "A. C. BETA" switch. This is the third circuit shifting switch, and as noted on the two lever switches, must be in the "RET" position except when set to perform its own specific test. These three controls may be used **one at a time** for measurements; however they are electrically interlocked, and each switch after use **must** be indexed back to the "RET" (Return) position before using any of the others. In making A. C. Beta measurements, this switch is first set to the "Set B" position, and the meter is then adjusted to the "Beta Set" mark on the dial as was explained in the section which described use of the "A. C. Beta" control. Once this adjustment has been made, the switch may be indexed to the 600 or 300 range and the A. C. Beta reading taken directly from the two top red arcs on the Collector Voltmeter. It should be remembered that a-c forward gain characteristics depend on d-c applied values. One of the very useful advantages of this analyzer is that variations in a-c Beta can be noted over a wide range of d-c conditions by observing the three meters operating simultaneously. Thus, h_{fe} can be plotted against I_C or I_B and/or E_C. NOTICE: This A. C. Beta range is operative only when the Collector Current switch is indexed to the .1A or lower ranges, and is thus limited to transistors which have a collector current requirement of .1 ampere or below.

Along the bottom left edge of the panel, the normal pilot lamp is located near the left hand edge. This lamp indicates as the analyzer is turned "On".

Next in line along the bottom edge of the panel, is the PNP-NPN lever switch. This switch controls the polarity of all potentials applied to the transistor or diode under test, and should be indexed correctly according to the manufacturer's data before making

any connections to the semi-conductor device.

Next, to the right is the "Collector Volts" switch. This determines the range of the meter directly above it and also does some power supply switching. The 3, 6 and 12 volt dc ranges of Collector Volts are equipped to deliver high current, as may be required for power transistors, and of course are also used for the signal types. These three positions, 3, 6, and 12 are the "high current" positions. In general, for power transistors the 3 or 6 volt meter range is adequate and should be used for those measurements requiring the full current rating of the transistor. Other measurements such as I_{CO}, I_{CEO}, Punch Thru, etc. where higher potentials may be required, call for comparatively low current requirements and can be made on 30, 60 and 120 volt ranges. Although the same high power rectifiers are used on all voltage ranges, lower power higher voltage transformer windings are used on the three top voltage ranges, and current should be limited to about 300 milliamperes. Hence, the operator should consider the three top voltage ranges, the "Low Current" ranges, and the three low voltage positions, ie. 3, 6 and 12 the "High Current" ranges.

CAUTION: When measuring high collector voltages (120V range) do not return to low voltage ranges until the charged filter capacitor discharges to a value that will not overload the meter.

CONFIGURATION—TEST INDEX

To the right of the "Collector Volts" switch is the "Common Emitter"—"Common Base" lever switch. This must be indexed for the configuration required. For the various measurements it should be indexed as follows:

COMMON EMITTER



1. DC Beta (h_{FE})
2. I_{CEO}
3. I_{CES}
4. I_{CER}
5. Base Current
6. Base Voltage
7. AC Beta h_{fe}
8. Saturation Voltage
9. Saturation Resistance
10. Any other grounded emitter measurement

COMMON BASE



1. Alpha (h_{FB})
2. I_{CO}
3. I_{CBS}
4. I_{CBR}
5. I_{EO}
6. I_{EBS}
7. I_{ECS}
8. I_{EBR}
9. Emitter Current
10. Emitter Voltage
11. Punch Thru
12. Any other grounded base measurement

The Collector control and Current range switch have been described before. These operate together for the adjustment of the Collector Current as described. To the far right the On-Off switch for power, and the fuse receptacle are apparent.

CONNECTIONS TO TRANSISTORS:

In the top center section of the panel are mounted two transistor sockets, directly above the central meter.

One of these is for the in-line types where the leads emerge from the bottom of the transistor in a straight line. (see Fig. 1) The second socket is for types such as where the leads are grouped in a circle or square on the bottom of the transistor. (see Fig. 2)

Nomenclature for socket terminals is as follows:

C—Collector
S—Ground or Shield
B—Base
E—Emitter



Fig. 1



Fig. 2

Aside from these two sockets, binding posts are available on the solid copper bar heat sink. The two large collector posts are mounted on and connect to the copper slab, which is insulated from the panel. The black insulated posts are from right to left, the base and the emitter connectors. The red insulated post is the b2 or tetrode connection. A socket for the most commonly used type of power transistor, that is those conforming to JEDEC outline TO-3 is built into the copper bar heat sink. Power transistors of this type have emitter and base pins extending from the bottom flange, and these fit into the two small holes between the two collector posts.



The base and also mounting flange, JEDEC No. E2-42 fits flush against the copper heat sink, and two clamps hold the transistor tight against the copper. This also provides the collector connection. Full collector current can be applied to these power types without excessive heat rise in the transistor. Examples of this type are the 2N1069, 2N1070 and 2N301.

Types such as the 2N101 may be mounted with screw stud on one of the collector posts, with the base and emitter leads either fastened directly or with extension clip leads furnished with the Model 3490.



Types such as the 2N441 and 2N1099 are mounted upside down with the slotted clamping plate fastened under the two collector posts.

CAUTION: IN CLAMPING THIS TYPE OF TRANSISTOR INTO PLACE, MAKE SURE THAT THE TWO PINS OF

THE TRANSISTOR WHICH ARE THE EMITTER AND BASE TERMINALS, ARE NOT TOUCHING THE CLAMP.



Where other types may be involved one of these combinations should provide sufficient thermal conductivity for any of the power types. If required, an external fitting may be used with leads fastened to the binding posts. A compartment is provided for such a device to the left rear of the connector block in the case.

On the signal types with long leads, it often saves time to connect the leads directly to the binding posts rather than insert them into the small sockets.



Where a transistor is equipped with a ground lead, this may be connected to the emitter post. On double base or tetrode units, the b2 terminal is connected to the red binding post. An example of this type is the 3N36.

THE HEAT SINK:

The solid slab of copper on which the binding posts are mounted forms with the two collector posts a substantial heat sink. This copper slab is insulated from the panel but is directly connected to the collector circuit. As mentioned under "Connections to Transistors", the copper and the brass nickel plated posts form the connection means to the power or signal transistors, aside from the 2 small sockets. When making connections to base, emitter or tetrode posts with bare leads, the operator should check to be sure these leads are spaced from the heat sink bar for insulation.

Buried in the copper on the under side is a socket for power transistors similar to the 2N457. Such transistors fit flush on the copper slab with the emitter and base wires on pins inserted in the two small holes marked "E" and "B" between the collector posts. Clamps fit under these two posts to hold the unit tight against the copper. Refer to the section entitled "Connections to Transistors".

THE INTERLOCKS:

To avoid overloads to transistors, instruments and analyzer components, circuit and switching interlocks are included in this device. On each position of the Input Switch, the impedance of the input supply is changed to provide as far as practical a constant current input supply regardless of emitter or base resistance. When the "Input Volts" lever switch is indexed to the "Punch Thru" position, the input (emitter) circuit is automatically opened regardless of the position of the Input Current switch. When the "I_{co}" lever switch is moved from its "RET" or Return position, the input circuit is automatically opened, regardless of the position of the Collector switch, thus affording instant entry of the sensitive I_{co} meter into the Collector circuit without any possibility of overload, even if the transistor under test had just previously been operating at 15 amperes collector current. The switching sequence is such that the emitter or base circuit is opened before the microammeter is connected into the collector circuit. Conversely, the microammeter is removed from the collector circuit **before** the input supply is connected restoring full collector current thru the collector ammeter when this switch is returned to the RET position.

THE WARNING LIGHTS:

Switching transients of very short duration can damage or destroy transistors, a condition not usual with vacuum tubes. Protective circuits have been designed into this analyzer to prevent such damaging conditions, with indicators in the form of warning lights for operator observation. Directly under the emitter binding post on the panel is a red light, which comes on when any potential appears from the input supply. This light is extinguished when the Input control is turned to the off or extreme counter clockwise position.

Under the collector posts is a second red light. This is illuminated when the Collector circuit is energized. This light is extinguished when the collector variac is turned to the off or extreme counter clockwise position. These two indicators are very important to the operation of the Analyzer. Whenever connections are changed at the transistor terminals or whenever the configuration switch and the PNP-NPN switch are touched, the lights should be out. Also, when changing current ranges on either Input or Collector switches it is advisable to have the lamps extinguished. This procedure protects the transistor at all times. The operator should get in the habit of returning both controls to the off position, and "running up from zero" watching the three meters.

PRECAUTION: DO NOT HANDLE CONNECTIONS OR INSERT OR REMOVE TRANSISTORS OR DIODES FROM SOCKETS WHEN LIGHTS ARE ON.

The third red warning light is of a smaller size. It is located to the upper right of the "Collector Volts" switch. This is connected to an interlock circuit between the Collector Current and Collector voltage switch. When the current switch is indexed to any current range **over** 300 milliamperes, and at the same time the collector voltage switch is indexed to any potential range from 30 to 120 volts, this light will be illuminated. For example, to assume the current switch is indexed to the 10 ampere position, and the collector voltage switch to the 60 volts position, this power requirement

of .6 kilowatts would overload the collector transformer and variac. The 10 ampere collector current can just as well be supplied on the 3, 6, or 12 volt range without such a high power requirement. By changing the "Collector Volts" switch to one of these lower voltage positions, the light will be extinguished. If on the other hand, a higher potential is required, the "Collector Volts" switch may be left in the 30, 60 or 120 volt position and the Collector Current switch indexed to the 100 ma, 30 ma, 10 ma or lower current position, where also the warning light will go out.

PRECAUTION: DO NOT PROCEED WITH TRANSISTOR OR DIODE MEASUREMENTS WHEN THIS LIGHT GLOWS RED.

MEASUREMENT OR TEST PROCEDURE

INFORMATION:

Locate the type number of the transistor in one of the data books available from manufacturers. For transistors note the characteristic, PNP or NPN. Set the lever switch accordingly.

Find the diagram showing the position of the connection leads and identify each one. Connect the leads to the binding posts or insert the unit in one of the sockets.

On Power types, clamp the unit to the heat sink to obtain good thermal contact using the brackets provided for this purpose. Check all connections.

TEST FOR SHORTED COLLECTOR:

Note the normal collector current listed and set the collector current switch accordingly. Index the Collector volts switch to the 6 volt position.

Be sure all warning lights are out, and the pilot is on. Check the maximum allowable collector voltage for the transistor to be sure it is at least six volts. Slowly rotate the collector control. The collector red warning light will first be illuminated and the collector voltmeter pointer will start up scale. Watch the collector current meter, and continue to increase the potential to 5 or 6 volts.

If the collector is free of shorts, the current meter will show little or no deflection. If this meter follows the voltmeter up scale with an appreciable deflection, then the collector is shorted to base or emitter, and the transistor is worthless. Put out the collector warning light by counter clockwise rotation of the Collector control, before disconnecting.

COLLECTOR JUNCTION LEAKAGE CURRENT MEASUREMENTS

MEASURING I_{CO}:

From the listed information, find the maximum allowed I_{CO} and the V_{CB} at which it is rated. Index the Collector Volts switch to cover the V_{CB} rating and the Configuration lever switch to Common Base. For example, a 2N111 shows a rated I_{CO} maximum of 5 microamperes at a Collector voltage of 12 volts.

Rotate the Collector control to the correct value of V_{CB} as read on the central meter. Shift the I_{CO} lever switch to the 60 ma position. This sets up the correct measurement circuitry for I_{CO} readings on the same meter that was reading collector volts. If, for example, a 2N174 is under test, the current reading can be taken on this range, since the I_{CO} max. is listed as 10 ma for this type. Read current on the black arc numbered zero to 6. If no deflection is apparent on any transistor under test on the 60 ma range, shift the lever to the 6 ma position. If again no deflection is noted, or if the pointer reads **below** the 10% scale deflection, shift the lever to the 600 microampere position, and read the current in microamperes. An additional range of 60 microamperes is available by push button for readings down to 1 microampere. For a transistor in normal operating condition, the current reading should be **less than** the max. rating in the manufacturer's specification. Referring back to the 2N111, the reading should be less than 5 microamperes at 12 volts. For the 2N174 also previously mentioned, the reading should be less than 10 milliamperes at 60 volts. Do not disturb the collector control when taking current readings, since this was previously set for the correct collector potential for the I_{CO} readings.

MEASURING I_{CEO}:

Refer to the listed information for the V_{CE} and note the potential. Index the Configuration lever switch to the Common Emitter position.

Rotate the Collector control to bring the pointer of the Collector Voltmeter to the correct V_{CE} reading. Shift the I_{CO}-I_{CEO} lever switch to the 60 ma position. This sets up the correct circuitry for I_{CEO} readings. Note the deflection on the center meter and read on the black arc marked zero to six, in terms of 60 ma full scale.

If the pointer does not deflect above the 10% mark, then this switch may be indexed to the 6 ma position, or lower ranges. In general, I_{CEO} readings are higher than I_{CO}. For a transistor, the reading should be **less than** the manufacturer's listed maximum.

Index the lever switch back to the RET position, and lower the collector voltage to zero.

MEASURING I_{CES}:

Refer to the listed information or otherwise determine the V_{CE} at which the measurement is to be made. Index the Configuration lever switch to the Common Emitter position. Be sure all warning lights are out.

Rotate the Input current switch to the Short position. Be sure the NPN-PNP switch is set correctly. Rotate the Collector control to bring the pointer of the Collector Voltmeter to the correct V_{CES} reading. Shift the I_{CO}-I_{CEO} lever switch to the 60 ma position. This sets up the correct circuitry for I_{CES} readings. As recorded under I_{CO} readings, shift the switch to a lower range and read the current on the black arc of the central meter marked zero to six, in terms of the corresponding lever switch position. For a normally operating transistor, the reading should be less than the manufacturer's listed maximum. Since the emitter and base are shorted this is also the value of I_{CBS}.

Index the lever switch back to the RET position, and lower the collector voltage to zero.

MEASURING I_{CER}:

Refer to the listed information or otherwise determine the V_{CE} at which the measurement is to be made. Index the Configuration lever switch to the Common Emitter position. Be sure all warning lights are out. Rotate the Input current switch to the Open position. Select the resistance ($_R$) value desired for the I_{CER} reading in accordance with the application of the transistor in its working circuit. A value of 10,000 ohms is quite common for this resistance or "$_R$", and may be used as a standard value if no further specific information is available. We will assume this value.

Be sure the NPN-PNP switch is set correctly. Clip a 10,000 ohm 1/4 or 1/2 watt resistor across the Emitter and Base binding posts. This can be made very convenient by making use of a General Radio type 274 or 274-MB connector. These binding posts have the 3/4" spacing that fits these connector plugs. Loosen the screws in the plug, insert the resistor leads and tighten them in place. This plug can then be inserted in the binding post tops even if the posts are in use for the transistor leads, or if the transistor under test is plugged into one of the sockets. The plug-in resistor thus available can be carried in the compartment of the analyzer as a useful accessory.

Rotate the Collector control to bring the pointer of the Collector Voltmeter to the correct V_{CER}. Shift the I_{CO}-I_{CEO} lever switch to a suitable current range position and read the I_{CER} on the black arc marked zero to six on the center meter.

MEASURING I_{CBR}:

This test is quite similar to the I_{CER} measurement, except that the configuration switch is indexed to the Common Base position, thus placing the resistance in the emitter circuit. Refer to the listed information or otherwise determine the I_{CB} at which the measurement is to be made. Index the configuration switch to the Common Base position. Be sure all warning lights are out. Index the Input current switch to the Open position. Select the resistance ($_R$) value desired for the I_{CBR} readings in accordance with the application of the transistor in its working circuit.

Be sure the NPN-PNP switch is set correctly and clip the chosen resistance across the Emitter and Base binding posts. Rotate the Collector control to bring the pointer of the Collector Voltmeter to the correct V_{CBR}. Shift the I_{CO}-I_{CEO} lever switch to a suitable current range position, and read the I_{CBR} on the black arc marked zero to six on the center meter.

EMITTER JUNCTION LEAKAGE CURRENT MEASUREMENTS

Any and all emitter or base input characteristics can be checked or studied with this analyzer other than those requiring high frequency apparatus, since independent full range continuous coverage of both voltage and current is available with multirange instruments. Care should be exercised in taking emitter or base readings where any appreciable potential is involved because the breakdown voltage of this part of any transistor is considerably lower than that of the collector circuit.

MEASURING I_{EO} (I_{EBO}):

Refer to the transistor manufacturer's information for the maximum V_{EBO} or V_{EB}. Index the configuration switch to Common Base. Be sure all warning lights are out. Index the Collector current switch to the Open position. Note the type of transistor, NPN or PNP. If the transistor under study is an NPN type, index the polarity switch to **PNP.** If the unit is a PNP type, index this switch to **NPN.**

If the V_{EBO} is 1 volt or less, index the Input Volts switch to the 1.2 volt range. If higher, index this switch to the 12 volt range. Set the Input current switch on the 300 ua range. Connect the transistor to the binding posts or plug into one of the sockets. Slowly rotate the Input control increasing the emitter potential to the value to be used in the circuit under design, or to some value less than the BV_{EBO}. The Input microammeter will read the sum of the leakage current and the emitter voltmeter current. Leave the Input control at its present setting, and shift the Input Volts lever switch back to the RET. position. This will subtract the voltmeter current from the leakage current. Read I_{EBO} in microamperes on the Input current meter. Return the Input control to zero extinguishing the warning light.

MEASURING I_{EBS} OR I_{ECS}:

Where low impedance output circuits are involved, the input or emitter leakage current with the collector short circuited is of more interest than I_{EBO}, and will be of a higher magnitude. Proceed as under instructions for measuring I_{EBO} in all respects except, index the Collector current switch to the "Short" position. Again index the polarity switch to PNP for the testing of NPN transistors, and to NPN for measurements on PNP types.

If BV_{EB} is 1 volt or less, index the Input Volts switch to the 1.2 volt range. Set the Input current switch to the 300 ua range. Be sure the warning lights are out. Connect the transistor to the binding posts or plug into one of the sockets. Slowly rotate the Input control increasing the emitter potential to the value to be used in the circuit under design, or to some value **less than** the BV_{EB}. The Input microammeter will read the sum of the leakage current and the emitter voltmeter current. To read I_{ECS}, leave the Input control at its present setting, and shift the Input Volts lever switch back to the RET. position. This will subtract the voltmeter current from the leakage current. Read I_{ECS} directly on the Input current meter. Return the Input control to zero extinguishing the warning light.

MEASURING I_{EBR}:

This is the input leakage current with the emitter reverse biased and a specific resistance, usually of fairly low value in the collector circuit. This measurement is of importance in circuit design particularly in class B amplifiers where the input characteristics must be known under definite load conditions.

Index the configuration switch to Common Base. Set the Collector current switch to the Open position. Index the polarity switch to PNP for the testing of NPN types, and to NPN for measurements on PNP types.

Index the Input Volts switch to the 1.2 volt range and the Input current switch to the 300 ua position. Be sure all warning lights are out. Connect a 1/4 or 1/2 watt resistor equal in value to the circuit requirement under design, between the Base and one of the collector posts. Plug in or connect the transistor. Rotate the Input control to apply the equivalent or required V_{EB} always less than the BV_{EB}. The Input microammeter will read the sum of the leakage current and the emitter voltmeter current. Leave the Input control at its setting, and shift the Input Volts lever switch back to the RET. position. This will subtract the voltmeter current. Read I_{EBR} directly on the Input current meter. This will fall somewhere between I_{EO} and I_{ECS}.

THE MEASUREMENT OF ALPHA, h_{FB}

Note the transistor type, NPN or PNP and set the lever switch accordingly. Index the Configuration switch to Common Base. Note the value of collector current I_C, at which the alpha rating is listed, and set the collector current switch to a suitable range for this current value. Set the input current switch to the same range. Be sure all warning lights are out.

Connect the transistor to the posts, or plug it into one of the sockets. Rotate the collector control slowly to a value of 5 or 6 volts as read on the collector voltmeter, unless some particular value is listed by the transistor manufacturer. Slowly rotate the input control watching the Input or Emitter Current. Bring this up to a value equal to the I_C listing: The collector current should follow closely the emitter current. Divide the collector current by the emitter current for the value of alpha. For example, the 2N319 is listed as having a minimum value of alpha of .94 at a collector current of 1 ma. In this case adjust the Input or Emitter Current to 1 ma. The collector current will be somewhere between .94 and .995 ma if the transistor is within limits. Use the 1 ma range on both current meters for this type. Other current ranges may be used depending on the transistor rating, or conditions under which the measurement is to be made, but in all cases where alpha is to be measured, the same Emitter and Collector Current ranges should be indexed on the current meters. In other words, if the 10 ma range is used for Emitter Current, then the 10 ma range should also be used for Collector Current, since alpha or h_F approaches unity. The Common Base configuration is always used for this measurement.

THE MEASUREMENT OF DC BETA, h_{FE}

Note the transistor type, NPN or PNP and set the lever switch accordingly. Index the Configuration switch to Common Emitter. Note the value of Collector current I_C, at which the beta average or minimum value is listed and set the collector current switch to a suitable range for this current value. Set the input current switch to a lower current range determined by the beta value listed for the transistor under test. If the expected beta is between 10 and 100 as most often will be the case, index the input current switch to the position that calls for one tenth (1/10) of the collector current switch reading. For example, if the collector current of a particular power transistor is listed as 10 amperes, index the collector current switch to the 10 ampere position, and the input current switch to the 1 ampere position. In general, the one tenth to one ratio is the best combination. Where higher values of beta are expected, a lower input current range may be used. Since h_{FE} is the ratio of collector current to input current, for best accuracy the lowest input range should be used that will deliver sufficient input current to drive the Collector current to its rated value.

For signal or low current transistors, note the normal operating collector voltage listed in the handbook or manufacturer's data sheet, and index the Collector Voltage switch to a suitable position for this potential. Be sure all three warning lights are out. Connect the transistor to the posts, or plug it into one of the sockets. Rotate the collector control to bring the center or collector voltmeter to the correct applied potential, using the right hand. With the left hand, slowly rotate the input control to apply base current .Both collector and input current meters will start up scale. Increase the input current until the collector current has reached the rated value, and correct the collector voltage, if necessary. Read the input current. Divide the Collector current by the input current for the measured value of h_{FE}. For example, a 2N1288 NPN switching transistor is shown in the handbook as having a minimum h_{FE} of 50, measured at a collector current of 10 milliamperes. The BV_{CE} is listed as 5 volts. To determine the dc beta of such a transistor, index the collector voltage switch to the six volt position and apply four volts to the collector (to stay under the 5 volt BV_{CE} rating). With the collector current switch indexed to the 10 ma position and the input current switch indexed to the .3 ma position, apply the input current necessary to bring the pointer of the collector current meter to the 10 ma reading. Divide 10 by the input current value. For a good transistor, this value should be more than 50.

The effect or influence of collector voltage on the value of dc beta can be studied by varying the potential on the collector and noting the current readings and their ratio. Curves may be plotted holding one of the values constant, or h_{FE} can be measured at any number of collector voltages, always holding the maximum V_{CE} below the BV_{CE} or breakdown potential.

In power or high current transistors, set the collector and input current switches as required and index the Collector Volts switch to the six volt position. Be sure all three warning lights are out. Clamp the transistor to the copper slab heat sink using one of the arrangements described in the section entitled "Connections to Transistors", page 5. Good contact with the heat sink must be maintained for both thermal transfer and electrical current flow. Check the NPN-PNP switch to be sure it is polarized correctly and set the Configuration switch to Common Emitter. Apply three volts to the collector using the right hand, and with the left slowly increase the input current. As the collector current increases, maintain the three volt collector potential by right hand rotation of the collector control. When rated collector current is flowing, read the input current and determine the current ratio for the h_{FE} value.

The three volt collector potential is used here on the power types simply to limit the power dissipation

in the collector supply. This same list could be made at 4 volts, or at 2 volts just as long as the collector voltage is above the saturation potential of the transistor. In general, three volts seems to be a generally acceptable figure. Where collector current above 10 amperes up to 20 amperes are involved, the operator should limit the time that the excess current over 10 amperes is applied to about thirty seconds. This avoids excessive temperature rise in the silicon rectifier supplying the collector current to the power transistor.

Again the influence of collector voltage on the value of dc beta can be explored by varying the collector voltage over a reasonable range and observing the collector to base current ratio.

MEASURING AC BETA, h_{fe}

In general, DC beta measurements correlate closely with AC measurements on power transistors, and therefore AC readings on power types are of little or no importance and are difficult to undertake. However, on low current or signal types h_{fe} tends to depart in value from h_{FE} and ac readings take on real significance. This is particularly true of transistors with collector currents under 100 milliamperes. This Analyzer is equipped with a 1 Kilocycle sine wave measuring circuit for full scale beta readings of 300 and 600 on the center meter. These scale arcs and the controls that are peculiar to this function carry red markings.

To measure h_{fe}, set up the controls normally for measuring h_{FE}, the dc beta, making sure that the collector current switch is indexed to a current range of 100 milliamperes, 30, 10, 3 or 1. All higher dc current ranges are automatically excluded from ac beta measurements. Apply the correct collector voltage and current by rotating the collector and input controls in accordance with the procedure for measuring dc beta. Index the AC Beta rotary switch to the Set B position. This removes the central meter from the collector voltage circuit, connects it into the ac beta circuit and energizes the 1 kc signal source. Rotate the Set Beta control to bring the pointer of the meter to the Set AC Beta mark directly above the 500 reading on the top red scale arc. This standardizes the ac beta circuit. Index the AC Beta switch to the 600 or 300 position, and read 600 or 300 full scale on the corresponding red arc.

Another example of the flexibility of this Analyzer is apparent here. The input current, collector current, or collector voltage may be shifted and curves plotted of ac beta against any of these variables. Also the dc conditions for maximum value of ac beta may be explored.

MEASURING SATURATION VOLTAGE AND RESISTANCE

The saturation voltage V_{CE} Sat. is measured in the Common Emitter configuration, and is the minimum voltage that will produce full collector current with the transistor current saturated. Index the NPN-PNP switch correctly. Refer to the transistor manufacturer's specification sheet for the input and collector current ratings for saturation tests. Index the input and collector current switches correctly. Be sure all warning lights are out. Index the Collector Voltage switch to the 3 volt position. Rotate the input control to increase the input or base current to the rated value as read on the input meter. Holding the base current constant, slowly apply a small amount of collector voltage until the collector current comes up to the listed value. Read the Saturation Voltage, V_{CE} Sat. on the central meter. For example, the type 2N457 transistor has a handbook listing of 5 amperes, collector current and 1 ampere base current, and under these conditions, the V_{CE} Sat. should not be over 1 volt. In the example we will assume a reading of .64 volts.

The Saturation Resistance can readily be determined by dividing the V_{CE} Sat. by I_C, the collector current. In the example above, V_{CE} Sat./I_C=.64/5=.128 ohm for Rsat.

MEASURING PUNCH THRU:

Punch Thru, sometimes referred to as "reach-thru", is a measure of the limit of emitter control with increasing collector potential. As the collector voltage V_{CB}, is increased, V_{EB}, the emitter to base will remain at a fixed low value in the millivolt region until the punch-thru voltage is reached. This effect of collector voltage on the emitter is caused by the spreading of the depletion layer into the emitter junction under the pressure of the increasing collector voltage. At the punch thru potential, the emitter to base voltage will suddenly start to increase linearly with increasing collector voltage, a signal that the punch-thru value has been exceeded. This is an important test since it will tell the operator the V_{CB} that should not be exceeded if normal input control is to be expected.

To make this measurement, first be sure all warning lights are out. Connect the transistor and index the NPN-PNP switch correctly. Set the configuration switch to "Common Base" position. Examine the manufacturer's data to see if the punch-thru potential is listed. Since this is a limit value, there is no great reason to run all transistors to the limit value, and therefore, if the transistor is being used at 6 volts or 9 volts in a particular application, measurements up to or somewhat above these values will be entirely sufficient.

Set the Collector Volts switch to the 12 volt position. Rotate the Collector control clockwise until the center meter indicates full scale, or 12 volts. There is a small index line on the collector control knob, that normally lines up with the panel index line at the off position. This line will now be to the right of the shaft center, at about the 3 o'clock position. Note where it is for the twelve volt reading. Return the control to zero so that the collector warning light is extinguished. Index the Input Volts switch to the Punch-Thru position. Rotate the collector control slowly clockwise to apply the collector voltage. As this potential is applied, the center meter will read "emitter floating potential" at a sensitivity of 1.2 volts full scale. For most small or low power transistors, the floating potential will come up to forty or fifty millivolts, and will hold steady as the collector voltage is increased. If there is no punch-thru, or in

other words the punch-thru potential is not reached when the collector control is rotated to the position noted before for 12 volts, then the meter reading will remain at approximately the original potential. In general, it should not rise over 10% scale or 120 millivolts during this test. If the punch-thru value is exceeded, the center meter will start up scale following the collector potential as the collector is rotated. When this occurs, the process should be repeated and the collector control stopped at the point where the meter reading or floating starts to increase. Then index the Input Volts switch back to the RET (return) position. The center meter will now read the Punch-Thru voltage of the transistor. It is assumed in making this test at 12 volts that the BV_{CB} or BV_{CE} potential listed for the transistor is greater than 12 volts. If not as is the case with the 2N523, this test must be run at 6 volts or lower. Under no conditions should a punch-thru test be made using collector potentials above the breakdown potential listing.

On power type transistors, the procedure is exactly the same. If the operator plans to use the transistor at 5 volts, a 6 volt punch-thru test is sufficient. If manufacturer's data is available, the test may be made at some other potential up to 12 volts V_C. In general the floating potential of the emitter or the power types may be somewhat higher than on signal units, but the test procedure is just the same.

The question may arise as to how to avoid avalanche breakdown in the transistor should the collector voltage be increased too rapidly when attempting to find punch-thru at collector potentials of 30 volts or above. When the Input Volts lever switch is indexed to the "Punch-Thru" position, a 390,000 ohm resistor is in series with the collector circuit, thus limiting the current and avoiding the avalanche breakdown.

With reasonable caution and following the procedure outlined above, punch-thru or reach-thru tests can be made without causing any damage to the transistors, and may be repeated when necessary. Occasionally a transistor will not exhibit a sharp break point, but will show increasing emitter floating potential at a fairly low Collector voltage, V_C. In such cases the punch-thru potential may be considered as the V_C that is required to raise the floating potential to 120 millivolts on the Punch-Thru range. This is equivalent to 5 microamperes of leakage current in the emitter circuit, the greater part of which is caused by collector voltage influence. Other values may be published as new transistors are announced, or more definite information is available on presently available types, but as a general rule to follow where more specific information is not available, the 120 millivolt limit will suffice.

FLOATING POTENTIAL:

The emitter floating potential at any collector voltage level below the punch of reach-thru value may be measured on the central meter. If the voltage on an open-circuited emitter is observed while the collector to base voltage is increased, it will remain within 500 millivolts or less of the base voltage until the punch-thru potential is reached. This test can be made by connecting the transistor with the warning lights extinguished, and a suitable collector voltage range selected, generally under 12 volts. The collector control should be rotated to apply the correct V_C and the index line position of the control noted. Return the collector voltage to zero and extinguish the collector warning light. Index the Input volts switch to the Punch-Thru position. Rotate the collector control back to the previously noted index line position, and read the floating potential on the central meter in terms of 1.2 volts full scale, or 1200 millivolts. The configuration switch must of course be in the Common Base position for these measurements.

TETRODE TRANSISTORS:

Tetrode transistors, primarily used in RF amplifiers up to 200 megacycles, and as mixers and oscillators require 3 separate and independent power supplies for test measurements. This analyzer is so equipped to handle these types. Examples of these transistors are the 3N36 and 3N37.

Tetrodes are equipped with a second base connection to which an interbase potential is applied. This reduces the base spreading resistance, R_B creating the equivalent effect of reduced cross section base region improving greatly the useful frequency of the unit. The so called "cross base bias" creates an electric field which compresses the active base region reducing the high frequency base resistance. Improvements of 10 to 1 are not uncommon in comparing the tetrode with a triode.

The tetrode or second base voltage supply is brought out to the red binding on the copper heat sink. The voltage control is directly to the left of the Collector Volts switch, and marked Tetrode Volts. This control is calibrated in volts from zero to 10.

Note the transistor type and index the NPN-PNP Switch accordingly. For alpha measurements, index the configuration switch to Common Base. Note the value of emitter current, and set both Input and Collector current switches to the same range. For the type 3N36 as an example, I_E is listed as—1.5 ma so the current switches should both be set to the 3 ma range. Note the listed collector voltage and set the Collector Volts switch to the required range. In the case of the referenced transistor, this is 5 volts. Be sure all warning lights are out. Plug in or otherwise connect the transistor, and apply the correct collector potential.

Rotate the Input control to bring the emitter current to the correct value and note the collector current Alpha under these conditions is the collector current divided by the emitter current. Slowly rotate the Tetrode Volts control up to the listed value. For the 3N36 this is —2 volts. Note the automatic gain control effect on the alpha value.

This same test may be repeated under Common Emitter conditions. Extinguish all warning lights and index the tetrode volts to zero. Set the Input and Collector current switches for beta measurements. In the case of the 3N36, a suitable collector range is 10 ma and input range is 300 ua. Apply a collector potential again of 5 volts and then apply input current to bring the collector current to 5 ma as a suggested figure. Note the dc beta. Slowly apply tetrode volts and observe the effect on the collector current and the resultant h_{FE}. For the 3N36 this should be limited to 2 volts.

AC Beta or h_{fe} may also be measured and the automatic gain control of the tetrode or second base observed. Return the tetrode volts to zero, switch to the Beta Set position and thence to the 300 range. Note the reading and then slowly increase the tetrode volts to the 2 volt potential.

Since all potentials are continuously variable and complete current and beta coverage is included in this Analyzer, curves showing one variable as a function of any of the others can be quickly and easily plotted. Efficient high frequency amplifiers with wide range coverage and smooth automatic gain control are quite practical with tetrode transistors.

TESTING DIODES AND RECTIFIERS

Information:

Locate the type number of the diode in one of the data books available from manufacturers. Note the polarity or direction of Current flow as usually indicated by an arrow. Note also the rated average forward current, the PIV, or peak inverse voltage, and the leakage current maximum at the peak inverse voltage.

On the larger power types with forward average current ratings over 3 amperes, it may be advisable to clamp the rectifier to the heavy brass posts on the copper slab if full current is to be applied for any considerable length of time. On many rectifiers in this class, the mounting stud can be inserted through the hole in the post.

Note the average forward current, and set the collector current switch to handle this value. Index the configuration switch to Common Base position, and the Collector Volts switch to 3 volts. Be sure all of the special purpose switches are in the RET position, and all warning lights are out.

TESTING DIODES:

Index the left hand lever switch to NPN. Connect the diode between the heavy collector and the base binding posts, with direction of normal or forward current flow from collector to base terminals. In the NPN switch position the collector terminals are positive. Rotate the collector control slowly watching the collector ammeter. If the forward resistance of the diode is low, this meter will start up scale. Keep increasing the collector current up to the listed average value. The full-load Forward Voltage Drop will be indicated by the center meter. For many germanium and silicon diodes in normal operating condition, this will be in the order of .3 to .5 volt.

If there is no apparent current, the diode may be open, or it may be connected in the reverse direction. To determine which is the case, return the collector control to zero extinguishing the warning light. Shift the polarity switch to the PNP position. Slowly rotate the collector control. If the collector current meter deflects up scale and can be brought to the average rated full load current, then the diode was reverse connected. On the other hand, if no current deflection is noted on either the NPN or PNP positions, then the diode is open circuited.

After determining the full load forward voltage drop, return the collector control to zero, extinguishing the warning light. Shift the polarity switch from the NPN position where the forward current was applied to the PNP position. Note the rated PIV, peak inverse voltage and index the collector voltage switch to encompass this value. For example, a 1N1093 computer diode is listed with a reverse working voltage of 15 volts. In this case set the Collector Voltage switch to 30 volts. Rotate the collector control to bring the pointer of the collector voltmeter up to the listed potential. There should be no deflection on the collector current meter. In the example just mentioned this meter would be operating on the .1 ampere range since the average anode current for this diode is 50 ma. If the current meter deflects up scale, then the diode has broken down in the reverse direction, and can not be used at the normal reverse potential. Normally there will be no appreciable current indication, and the I_{CO}-I_{CEO} switch may be shifted down through the current ranges until a readable deflection is noted. This is the leakage current at the reverse working voltage, sometimes referred to as the reverse current. For the 1N1093 this should be less than 75 ua at 15 volts. Some manufacturers list other reverse current values at potentials considerably below the peak inverse voltage or continuous reverse working voltage. For example, a second value of 25 ua at 5 volts is specified for this same diode. In all cases, the leakage or reverse current should be **less than** the manufacturer's maximum. Return the I_{CO}-I_{CEO} switch to the RET position before proceeding with any other steps. Return the collector voltage to zero extinguishing the collector warning lamp.

Where peak inverse or reverse working potentials are encountered higher than 120 volts, measurements will have to be confined to this potential since this is the maximum available in the Analyzer. Under some low line voltage conditions, this may be closer to 110 V.

There are many types, kinds and varieties of diodes where the operator will be interested in additional tests. In some instances curve data may be valuable to pin point break points in the characteristics. By studying the description of the controls and following the test procedures already described he can work out for himself other procedures and measurements that will be of considerable value in the study of semiconductors.

TESTING RECTIFIERS:

Rectifiers are checked, measured, and tested in about the same manner as diodes, except that in general the working current values are higher. As mentioned in the second paragraph under "Information", heat sink considerations must be taken into account.

Proceed as under Diodes with the Collector current switch set to the correct position for the rated average anode current. Determine which is the cathode and which the anode of the rectifier. On heat sink types where the average anode current is of sufficient amplitude to require conduction cooling of the junction, fasten or clamp the stud or heat sink end to the large collector posts, or against the copper slab. If this happens to be the anode, index the polarity switch to the NPN position for forward conduction tests. On types where this is the cathode, index to the PNP position.

On silicon and germanium rectifiers, the collector voltage switch should be indexed to the 3 volt position for forward current tests. In general the threshold volt-

age is less than half a volt. This range provides the most accurate reading of full load forward voltage drop. Many of these rectifiers and diodes are listed by their manufacturers for minimum and maximum forward current at 1 volt. This can be measured quickly and easily using the 3 volt collector position. An example would be the 1N484B High Temperature Diode with a minimum forward current rating of 100 milliamperes at 1 volt.

SELENIUM RECTIFIERS:

Selenium rectifiers do not have as low a forward resistance as germanium and silicon types. Where single discs are being tested or measured, the 3 volt range should still be adequate for full current forward voltage drop. Occasionally it may be necessary to use the 6 volt or the 12 volt range for collector potential. This is especially true of the rated 120 volt stacks which may have a full load forward voltage drop of 6 or 7 volts. In all other respects, the test procedure is the same as for the silicon types.

Copper oxide rectifiers are similar in forward characteristics to selenium. Care should be exercised in testing these units to be sure the peak inverse voltage is not exceeded when measuring reverse current.

PRECAUTION—WHEN TESTING ALL SEMI-CONDUCTORS

The I_{CO}-I_{CEO} switch should be indexed into any current range for a limited interval only, and must be shifted back to the RET position as fast as a current reading can be taken. It is advisable to keep the fingers of the right hand on this switch all the time that it is in use.

BASIC TRANSISTOR TEST CIRCUITS



WIRING DIAGRAM, 3490



REPLACEABLE PARTS LIST, 3490

Ref. No.	Part No.	Req.	Part Name	Description
M1	52-1525	1	Instrument	420,200 Ua, Adjust to 500 ohms
M2	52-1526	1	Instrument	420, 50 Ua, Adjust to 2000 ohms
M3	52-1527	1	Instrument	420,200 Ua, Adjust to 500 ohms
S1	22A-379	1	Switch	Lever, PNP-NPN
S2	22A-372	1	Switch	Lever, Common Emitter-Common Base
S3	22A-419	1	Switch	Collector Current
S4	22A-370	1	Switch	Input Current
S5	22A-420	1	Switch	Lever, Input Volts
S6	22A-421	1	Switch	Lever, I_{CO}-I_{CEO}
S7	22A-367	1	Switch	Collector Volts
S8	22A-374	1	Switch	Push Button
S9, 10	22A-375	2	Switch	No Detent Single Circuit, Warning Lights
S11	22-116	1	Switch	Toggle, On-Off
S12	22A-377	1	Switch	A. C. Beta
	22-412	1	Switch Assembly	Collector Current, With Resistors
	22-413	1	Switch Assembly	Input Current, With Resistors
	22-414	1	Switch Assembly	Input Volts, With Resistors
	22-415	1	Switch Assembly	I_{CO}-I_{CEO}, with Resistors
	22-416	1	Switch Assembly	Collector Volts, With Resistors
	22-417	1	Switch Assembly	Push Button, With Resistors
	22-418	1	Switch Assembly	A. C. Beta, With Resistors
T1	23A-159	1	Transformer	1 KC Oscillator Coil
T2	23A-122	1	Transformer	A. C. Beta
T3	23A-158	1	Transformer	Collector Current, 100 W
T4	23A-115	1	Transformer	Input Current, 25 Watt
T5	23A-124	1	Transformer	Variable Powerstat
Q1	2455-232	1	Socket	Power Transistor
Q2	2455-223	1	Socket	4 Pin Transistor, Rectangular, With Retainer Ring
Q3	2455-221	1	Socket	4 Pin Transistor, Round, With Retainer Ring
F1	3207-20	1	Fuse	3 Amp, 8 Ag
	3207-19	1	Fuse Holder	
L1, 2, 3, 4	67-99	4	Lamps	Pilot, 6 Volt, (Bayonet Base)
X1, 2, 3	2250-47	3	Diodes	1N38A
X4, 5	2250-45	2	Rectifier	10 Amp Silicon
X6	2250-46	1	Rectifier	Full Wave
X7	2250-40	1	Rectifier	Half Wave
C1	43-238	1	Capacitor	200 MFD, 150 WVDC
C2	43-125	1	Capacitor	.02 MFD, 600 WVDC, Ceramic Disc
C3	43-248	1	Capacitor	1000 MFD, 15 WVDC
C4	43-240	1	Capacitor	20 MFD, 150 WVDC
C5	43-108	1	Capacitor	.5 MFD, 200 WVDC
C6	43-245	1	Capacitor	6000 MFD, 15 WVDC
Z1	127-1	1	Transistor	2N591
R1	T-15-1447	1	Resistor	6.5K ohm ±1%, ½ Watt, Film Type
R2	T-15-1179	1	Resistor	1500 ohm ±1%, ½ Watt, Film Type
R3	T-15-1487	1	Resistor	47K ohm ±5%, ½ Watt, Composition
R4	T-2601-½-1K	1	Resistor	1K ohm ±10%, ½ Watt, Composition
R5	T-15-2577	1	Resistor	2.2K ohm ±1%, ½ Watt, Film Type
R6	T-15-3339	1	Resistor	700 ohm ±1%, ½ Watt, Film Type
R7	T-15-2306	1	Resistor	5.6K ohm ±10%, 2 Watt, Composition
R8	16-129	1	Resistor	175 ohm Variable
R9, 63	16-122	2	Resistor	750 ohm Variable
R10	90-650	1	Shunt Assem.	10 Amp
R11	90-651	1	Shunt Assem.	30 Amp
R12	T-15-1009	1	Resistor	5K ohm ±1%, ½ Watt, Film Type
R13	15-3343	1	Resistor	22K ohm ±1%, ½ Watt, Film Type
R14	15-2922	1	Resistor	15K ohm ±1%, ½ Watt, Film Type
R15	T-2601-½-1000	1	Resistor	1K ohm ±10%, ½ Watt, Composition
R16	T-16-90	1	Resistor	3500 ohm Variable
R17 & 64	15-3456	2	Resistor	3A .14 ohm ±½%, Wirewound .15
R18	15-3215	1	Resistor	1A .2 ohm ±½%, Wirewound .208
R19 & 27	15-3381	2	Resistor	.3A .7 ohm ±½%, Wirewound 1.5
R20 & 28	15-3043	2	Resistor	.1A 2 ohm ±½%, Wirewound 2.0
R21 & 29	T-15-2058	2	Resistor	30mA 7 ohm ±½%, Wirewound
R22 & 30	15-3045	2	Resistor	10mA 20 ohm ±½%, Wirewound

3.7+3.16+.15=7.01

2,402

REPLACEABLE PARTS LIST, 3490

Ref. No.	Part No.	Req.	Part Name	Description
R23 & 31	15-3363	2	Resistor	70 ohm ±½%, Wirewound
R24 & 32	T-15-3468	2	Resistor	200 ohm ±½%, ½ Watt, Film Type
R25 & 33	15-3339	2	Resistor	700 ohm ±½%, ½ Watt, Film Type
R26	15-3382	1	Resistor	.3 ohm ±½%, Wirewound
R34	15-3409	1	Resistor	5 ohm ±5%, 5 Watt
R35	T-15-3410	1	Resistor	15 ohm ±5%, 5 Watt
R36	15-3344	1	Resistor	60 ohm ±1%, 1 Watt, Film Type
R37	15-3336	1	Resistor	220 ohm ±5%, ½ Watt, Composition
R38	T-2601-½-680	1	Resistor	680 ohm ±10%, ½ Watt, Composition
R39	T-2601-½-2700	1	Resistor	2.7K ohm ±10%, ½ Watt, Composition
R40	16-135	1	Rheostat	25 ohm, 25 Watt
R41	16-102	1	Rheostat	2.5K ohm, ±10%, 2 Watt
R42	15-3337	1	Resistor	6.8K ohm ±10%, ½ Watt, Composition
R43	T-2601-½-27K	1	Resistor	27K ohm ±10%, ½ Watt, Composition
R44	15-3342	1	Resistor	216K ohm ±½%, ½ Watt, Film Type
R45	15-3343	1	Resistor	22K ohm ±½%, ½ Watt, Film Type
R46	15-2059	1	Resistor	10 ohm ±½%, ½ Watt, Film Type
R47	15-3340	1	Resistor	90 ohm ±½%, ½ Watt, Film Type
R48	15-3469	1	Resistor	900 ohm ±½%, ½ Watt, Film Type
R49	T-2601-1-1K	1	Resistor	1K ohm ±10%, 1 Watt, Composition
R50	15-3341	1	Resistor	58K ohm ±½%, ½ Watt, Film Type
R51	T-15-3471	1	Resistor	60K ohm ±½%, ½ Watt, Film Type
R52	T-15-3472	1	Resistor	120K ohm ±½%, ½ Watt, Film Type
R53	T-15-3473	1	Resistor	360K ohm ±½%, ½ Watt, Film Type
R54	T-15-3474	1	Resistor	600K ohm ±½%, ½ Watt, Film Type
R55	T-15-2438	1	Resistor	1.2 Meg ohm ±1%, ½ Watt, Film Type
R56 & 66	T-15-3411	2	Resistor	20 ohm ±5%, 1 Watt, Composition
R57	T-15-3470	1	Resistor	9K ohm ±½%, ½ Watt, Film Type
R58	T-15-2442	1	Resistor	100K ohm ±½%, ½ Watt, Film Type
R59	T-15-1061	1	Resistor	200K ohm ±½%, ½ Watt, Film Type
R60	15-3085	1	Resistor	22K ohm ±5%, ½ Watt, Composition
R61	3995	1	Resistor	2K ohm ±5%, ½ Watt, Composition
R62	T-16-117	1	Resistor	500 ohm ±20%, Variable
R65	T-15-2347	1	Resistor	390K ohm ±10%, ½ Watt, Composition
R67	T-15-1425	1	Resistor	160 ohm ±10%, 1 Watt
	79A-176	1	Lead Assembly	
	34-77	5	Knobs	Pointer Type
	34-76	2	Knobs	Skirted
	34-78	1	Knob	Round
	T-34A-73	4	Knob	Lever
	2566-51	1	Line Cord	7 ft. long
	10A-1642	1	Case	Carrying

TRIPLETT WARRANTY AND CONDITIONS OF SALE

The Triplett Electrical Instrument Company warrants instruments manufactured by it to be free from defective material or factory workmanship and agrees to repair or replace such instruments which under normal use and service, disclose the defect to be the fault of our manufacturing. Our obligation under this warranty is limited to repairing or replacing any instrument or test equipment which proves to be defective, when returned to us transportation prepaid, within ninety (90) days from the date of original purchase.

This warranty does not apply to any of our products which have been repaired or altered by unauthorized persons or service stations in any way so as, in our judgment, to injure their stability or reiability or which have been subject to misuse, negligence, or accident, or which have had the serial number altered, effaced, or removed. Neither does this warranty apply to any of our products which have been connected, installed, or adjusted otherwise than in accordance with the instructions furnished by us. Accessories including all vacuum tubes and batteries not of our manufacture used with this product are not covered by this warranty.

The Triplett Electrical Instrument Company reserves the right to discontinue models at any time, or change specifications or design, without notice and without incurring any obligation.

Upon acceptance of the material covered by this invoice the purchaser agrees to assume all liability for any damage and bodily injury which may result from the use or misuse of the material by the purchaser, his employees, or others, and that The Triplett Electrical Instrument Company shall incur no liability for direct or consequential damage of any kind.

Parts will be made available for a maximum period of five (5) years after the manufacture of this equipment has been discontinued. Parts include all materials, charts, instructions, diagrams, accessories, et cetera, which were furnished in the standard or special models.

This warranty and conditions of sale are in lieu of all others expressed or implied and no representative or person is authorized to assume for us any other liability in connection with the sale of our products.

The Triplett Electrical Instrument Company

Printed in U. S. A. Bluffton, Ohio Part No. 84-118-060162-3

MP MASTER
MA26153